

# Compliance Report Of Technical Committee For Electrical and Electronics Department Labs Equipment

## Lab: Electrical Circuit Lab

S.No	Equipment Name	Intermark	Muslim Scientific Corporation	Rohde & Schwarz
1	Digital Oscilloscope	Not Compliant due to ISO-14001 standard	Not Compliant due to ISO-14001 standard	Compliant (RTC 1002)
2	Digital Multimeter	Not Compliant due to low over-current protection	Not Compliant due to low over-current protection	Not Compliant due to low over-current protection
3	Dual Power Supply	Not Compliant due to ISO-14001 standard	Not Compliant due to ISO-14001 standard	Compliant (HMC 8042)
4	Function Generator	Not Compliant due to Wavelength and other specifications	Not Compliant due to Wavelength and other specifications	Not Compliant due to wavelength and other specifications




**Lab: Electronics Devices and CKT Lab**

S.No	Equipment Name	Intermark	Muslim Scientific Corporation	Rohde & Schwarz
1	Digital Oscilloscope	Not Compliant due to ISO-14001 standard	Not Compliant due to ISO-14001 standard	Compliant (RTC 1002)
2	Digital Multimeter	Not Compliant due to low over-current protection	Not Compliant due to low over-current protection	Not Compliant due to low over-current protection
3	Dual Power Supply	Not Compliant due to ISO-14001 standard	Not Compliant due to ISO-14001 standard	Compliant (HMC 8042)
4	Function Generator	Not Compliant due to Wavelength and other specifications	Not Compliant due to Wavelength and other specifications	Not Compliant due to wavelength and other specifications

**Recommended By Technical Committee Members:**

1. Engr. ~~Amjad~~ Hamza (Technical Member)



2. Engr. Saad Ijaz Majid (Technical Member)



3. Engr. Saadia Tabassum (Technical Member)



**Compliance Report Forwarded and Checked by: Dr. Haider Ali**

